



Edition 1.1 2024-10 CONSOLIDATED VERSION

INTERNATIONAL STANDARD



Electrostatics -

Part 6-1: Electrostatic control for healthcare - General requirements for facilities Electrostatic control in healthcare, commercial and public facilities - Healthcare

INTERNATIONAL ELECTROTECHNICAL COMMISSION

ICS 17.220.99, 29.020, 11.020.99

ISBN 978-2-8322-9913-5

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

F	OREWO	PRD	3
۱N	NTRODU	JCTION	5
1	Scop	pe	6
2	Norr	native references	6
3	Tern	ns and definitions	7
4	Electrostatic hazards		
	4.1	General	
	4.2	ESD effects on equipment	
	4.3	Contamination caused by ESA	
	4.4	Ignition of flammable substances	
	4.5	Electrostatic shock to people	10
5	Elec	trostatic control requirements	10
	5.1	General	10
	5.2	Medical procedures	10
	5.3	Medical locations	10
	5.3.	Classification by groups	10
	5.3.2	Unclassified rooms	11
	5.3.3	Group 0 – Electrostatic control recommended	11
	5.3.4	Group 1 – Electrostatic control recommended conditionally required	11
	5.3.5	Group 2 – Electrostatic control required	12
	5.4 Service and maintenance		
	5.5	Administrative requirements and recommendations	12
	5.5.	Designing facilities	12
	5.5.2	P Operational responsibility	12
	5.5.3	Qualification and verification	12
	5.6	Technical requirements	
	5.6.	,	
	5.6.2		
	5.6.3		
	5.7	Packaging, containers and other electrostatic control items	
A	nnex A	(normative) Test methods for low charging textiles	
	A.1	Test methods for clothing and upholstery	
	A.2	Test methods for bedding, curtains, and surgical drapes	17
A	nnex B	(informative) Ionization and other considerations	21
В	ibliogra	phy	22
		1 – Example of test equipment set up for measuring body voltage when item of bedding or surgical drape from person wearing reference clothing	18
		2 – Example of test equipment set up for measuring body voltage when item of bedding or surgical drape from bed or examination/operating table	19
		3 – Example of test equipment set up for measuring body voltage on two nen removing item of bedding or surgical drape	20
т	ahle 1 _	Summary of electrostatic control methods for specified locations	11

INTERNATIONAL ELECTROTECHNICAL COMMISSION

ELECTROSTATICS -

Part 6-1: Electrostatic control for healthcare – General requirements for facilities

Electrostatic control in healthcare, commercial and public facilities – Healthcare

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at https://patents.iec.ch. IEC shall not be held responsible for identifying any or all such patent rights.

This consolidated version of the official IEC Standard and its amendment has been prepared for user convenience.

IEC 61340-6-1 edition 1.1 contains the first edition (2018-09) [documents 101/566/FDIS and 101/570/RVD] and its amendment 1 (2024-10) [documents 101/713/FDIS and 101/720/RVD].

In this Redline version, a vertical line in the margin shows where the technical content is modified by amendment 1. Additions are in green text, deletions are in strikethrough

red text. A separate Final version with all changes accepted is available in this publication.

International Standard IEC 61340-6-1 has been prepared by IEC technical committee 101: Electrostatics.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts in the IEC 61340 series, published under the general title *Electrostatics*, can be found on the IEC website.

The committee has decided that the contents of this document and its amendment will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

IMPORTANT - The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

© IEC 2024

INTRODUCTION

Static electricity can be the source of several hazards to patients, staff and equipment in healthcare facilities. Such hazards include:

- electromagnetic disturbance or electrostatic discharge (ESD) disruption or damage to medical instrumentation and data processing equipment;
- damage to ESD susceptible electronic components and assemblies during service and maintenance:
- electrostatic attraction (ESA) and contamination;
- ignition of flammable gases, liquids and other materials, and
- electrostatic shocks to people.

Adequate electrostatic control can eliminate these hazards, or at least reduce residual risk to tolerable levels.

ELECTROSTATICS -

– 6 –

Part 6-1: Electrostatic control for healthcare –
General requirements for facilities
Electrostatic control in healthcare,
commercial and public facilities – Healthcare

1 Scope

This part of IEC 61340 applies to facilities that provide healthcare including hospitals, care centres and clinics.

This document provides technical requirements and recommendations for controlling electrostatic phenomena in healthcare facilities, which includes requirements for equipment, materials, and products used to control static electricity.

The requirements of this document do not apply to medical electrical equipment specified in IEC 60601-1 [1] ¹ and in vitro diagnostic (IVD) medical equipment specified in IEC 61010-2-101 [2].

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60364-7-710, Electrical installations of buildings Low-voltage electrical installations – Part 7-710: Requirements for special installations or locations – Medical locations

IEC TR 61340-1, Electrostatics – Part 1: Electrostatic phenomena – Principles and measurements

IEC 61340-2-1, Electrostatics – Part 2-1: Measurement methods – Ability of materials and products to dissipate static electric charge

IEC 61340-2-3, Electrostatics – Part 2-3: Methods of test for determining the resistance and resistivity of solid materials used to avoid electrostatic charge accumulation

IEC 61340-4-1, Electrostatics – Part 4-1: Standard test methods for specific applications – Electrical resistance of floor coverings and installed floors

IEC TS 61340-4-2:2013, Electrostatics – Part 4-2: Standard test methods for specific applications – Electrostatic properties of garments

IEC 61340-4-3, Electrostatics – Part 4-3: Standard test methods for specific applications – Footwear

Numbers in square brackets refer to the bibliography.

IEC 61340-4-5, Electrostatics — Part 4-5: Standard test methods for specific applications — Methods for characterizing the electrostatic protection of footwear and flooring in combination with a person

IEC 61340-5-1, Electrostatics – Part 5-1: Protection of electronic devices from electrostatic phenomena – General requirements

ISO 18080-2, Textiles – Test methods for evaluating the electrostatic propensity of fabrics – Part 2: Test method using rotary mechanical friction

ISO 18080-3, Textiles – Test methods for evaluating the electrostatic propensity of fabrics – Part 3: Test method using manual friction

ISO 18080-4, Textiles – Test methods for evaluating the electrostatic propensity of fabrics – Part 4: Test method using horizontal mechanical friction

ISO 20344, Personal protective equipment – Test methods for footwear

CONTENTS

FC	REWO	RD	3
IN	TRODU	CTION	5
1	Scop	e	6
2	Norm	ative references	6
3	Term	s and definitions	7
4	Elect	rostatic hazards	8
	4.1	General	
	4.2	ESD effects on equipment	
	4.3	Contamination caused by ESA	
	4.4	Ignition of flammable substances	
	4.5	Electrostatic shock to people	10
5	Elect	rostatic control requirements	10
	5.1	General	10
	5.2	Medical procedures	10
	5.3	Medical locations	10
	5.3.1	Classification by groups	10
	5.3.2	Unclassified rooms	11
	5.3.3	•	
	5.3.4	,	
	5.3.5	·	
	5.4	Service and maintenance	
	5.5	Administrative requirements and recommendations	
	5.5.1	Designing facilities	
	5.5.2	, ,	
	5.5.3		
	5.6	Technical requirements	
	5.6.1	Electrical safety	
	5.6.2		
	5.6.3		
۸ ۵	5.7	Packaging, containers and other electrostatic control items	
ΑΠ	`	,	
	A.1	Test methods for clothing and upholstery	
۸	A.2	Test methods for bedding, curtains, and surgical drapes	
		informative) Ionization and other considerations	
BI	ollograp	hy	22
rei Fig	moving gure A.2	I – Example of test equipment set up for measuring body voltage when item of bedding or surgical drape from person wearing reference clothing 2 – Example of test equipment set up for measuring body voltage when item of bedding or surgical drape from bed or examination/operating table	
	_	B – Example of test equipment set up for measuring body voltage on two	
		nen removing item of bedding or surgical drape	20
Та	ble 1 –	Summary of electrostatic control methods for specified locations	11

INTERNATIONAL ELECTROTECHNICAL COMMISSION

ELECTROSTATICS -

Part 6-1: Electrostatic control in healthcare, commercial and public facilities – Healthcare

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at https://patents.iec.ch. IEC shall not be held responsible for identifying any or all such patent rights.

This consolidated version of the official IEC Standard and its amendment has been prepared for user convenience.

IEC 61340-6-1 edition 1.1 contains the first edition (2018-09) [documents 101/566/FDIS and 101/570/RVD] and its amendment 1 (2024-10) [documents 101/713/FDIS and 101/720/RVD].

This Final version does not show where the technical content is modified by amendment 1. A separate Redline version with all changes highlighted is available in this publication.

Electrostatics.

International Standard IEC 61340-6-1 has been prepared by IEC technical committee 101:

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts in the IEC 61340 series, published under the general title *Electrostatics*, can be found on the IEC website.

The committee has decided that the contents of this document and its amendment will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- · withdrawn, or
- revised.

© IEC 2024

INTRODUCTION

Static electricity can be the source of several hazards to patients, staff and equipment in healthcare facilities. Such hazards include:

- electromagnetic disturbance or electrostatic discharge (ESD) disruption or damage to medical instrumentation and data processing equipment;
- damage to ESD susceptible electronic components and assemblies during service and maintenance:
- electrostatic attraction (ESA) and contamination;
- ignition of flammable gases, liquids and other materials, and
- electrostatic shocks to people.

Adequate electrostatic control can eliminate these hazards, or at least reduce residual risk to tolerable levels.

ELECTROSTATICS -

– 6 –

Part 6-1: Electrostatic control in healthcare, commercial and public facilities – Healthcare

1 Scope

This part of IEC 61340 applies to facilities that provide healthcare including hospitals, care centres and clinics.

This document provides technical requirements and recommendations for controlling electrostatic phenomena in healthcare facilities, which includes requirements for equipment, materials, and products used to control static electricity.

The requirements of this document do not apply to medical electrical equipment specified in IEC 60601-1 [1] ¹ and in vitro diagnostic (IVD) medical equipment specified in IEC 61010-2-101 [2].

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60364-7-710, Low-voltage electrical installations – Part 7-710: Requirements for special installations or locations – Medical locations

IEC TR 61340-1, Electrostatics – Part 1: Electrostatic phenomena – Principles and measurements

IEC 61340-2-1, Electrostatics – Part 2-1: Measurement methods – Ability of materials and products to dissipate static electric charge

IEC 61340-2-3, Electrostatics – Part 2-3: Methods of test for determining the resistance and resistivity of solid materials used to avoid electrostatic charge accumulation

IEC 61340-4-1, Electrostatics – Part 4-1: Standard test methods for specific applications – Electrical resistance of floor coverings and installed floors

IEC TS 61340-4-2:2013, Electrostatics – Part 4-2: Standard test methods for specific applications – Electrostatic properties of garments

IEC 61340-4-3, Electrostatics – Part 4-3: Standard test methods for specific applications – Footwear

IEC 61340-4-5, Electrostatics – Part 4-5: Standard test methods for specific applications – Methods for characterizing the electrostatic protection of footwear and flooring in combination with a person

¹ Numbers in square brackets refer to the bibliography.

© IEC 2024

IEC 61340-5-1, Electrostatics – Part 5-1: Protection of electronic devices from electrostatic phenomena – General requirements

ISO 18080-2, Textiles – Test methods for evaluating the electrostatic propensity of fabrics – Part 2: Test method using rotary mechanical friction

ISO 18080-3, Textiles – Test methods for evaluating the electrostatic propensity of fabrics – Part 3: Test method using manual friction

ISO 18080-4, Textiles – Test methods for evaluating the electrostatic propensity of fabrics – Part 4: Test method using horizontal mechanical friction

ISO 20344, Personal protective equipment – Test methods for footwear