



IEC 62047-48

Edition 1.0 2024-06

# INTERNATIONAL STANDARD



---

**Semiconductor devices – Micro-electromechanical devices –  
Part 48: Test method for determining solution concentration by optical  
absorption using MEMS fluidic device**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

---

ICS 31.080.99

ISBN 978-2-8322-8930-3

**Warning! Make sure that you obtained this publication from an authorized distributor.**

## CONTENTS

FOREWORD.....	3
INTRODUCTION.....	5
1 Scope.....	6
2 Normative references .....	6
3 Terms and definitions .....	6
4 Test method .....	6
4.1 General.....	6
4.1.1 Principle of the absorption method.....	6
4.1.2 Molar absorption coefficient, $a$ .....	8
4.1.3 Optical path length, $b$ .....	8
4.1.4 Absorbance, $A$ .....	8
4.1.5 Surface temperature of MEMS fluidic device.....	8
4.1.6 Molar concentration, $c$ .....	8
5 Test specimen .....	9
5.1.1 Test specimen .....	9
5.1.2 Configuration of concentration measuring unit for test specimen.....	9
5.1.3 Layer stack of the MEMS fluidic device.....	9
5.1.4 Size and shape of the micro-channel .....	10
5.1.5 Configuration of solution concentration measuring system.....	11
5.2 Test procedure.....	13
6 Data analysis.....	14
6.1 Calculating results .....	14
6.2 Test report .....	14
Bibliography.....	16
Figure 1 – Schematic drawing of micro components in a MEMS fluidic device (top view) .....	5
Figure 2 – Optical absorption method for determining the concentration of the solution using a conventional cuvette (side view) .....	7
Figure 3 – Schematic graph with the x-axis as the concentration, $c$ and the y-axis as absorbance $A$ .....	8
Figure 4 – Schematic drawing of a concentration measuring unit for a test specimen based on a MEMS fluidic device .....	9
Figure 5 – Schematic structure and process for preparation of MEMS fluidic device .....	10
Figure 6 – Configurations and dimensions for measuring unit of MEMS fluidic device .....	11
Figure 7 – Schematic drawing for a MEMS fluidic device layout with microfluidic concentration measuring unit and additional MEMS components on a single device .....	12
Figure 8 – Schematic drawing for stand-alone measuring unit connected with another MEMS fluidic device through tubing .....	13
Table 1 – Test unit for determining solution concentration using MEMS fluidic device.....	10
Table 2 – Table for test report for determining solution concentration by absorption method in MEMS fluidic device .....	15

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

SEMICONDUCTOR DEVICES –  
MICRO-ELECTROMECHANICAL DEVICES –

**Part 48: Test method for determining solution concentration  
by optical absorption using MEMS fluidic device**

## FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 62047-48 has been prepared by subcommittee 47F: Micro-electromechanical systems, of IEC technical committee 47: Semiconductor devices. It is an International Standard.

The text of this International Standard is based on the following documents:

Draft	Report on voting
47F/466/FDIS	47F/472/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at [www.iec.ch/members\\_experts/refdocs](http://www.iec.ch/members_experts/refdocs). The main document types developed by IEC are described in greater detail at [www.iec.ch/publications](http://www.iec.ch/publications).

A list of all parts in the IEC 62047 series, published under the general title *Semiconductor devices – Micro-electromechanical devices*, can be found on the IEC website.


The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under [webstore.iec.ch](http://webstore.iec.ch) in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

**IMPORTANT – The "colour inside" logo on the cover page of this document indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.**

## INTRODUCTION

A MEMS fluidic device is one of the key devices in MEMS technologies, including bio-MEMS, chemical MEMS, and micro TAS (total analytical system). A MEMS fluidic device, in general, consists of several micro components such as inlet ports for injection of a filtered sample and reagents to induce a sample to have the optical absorption at specific wavelength, a microfluidic mixer for physical mixing, a micro-reactor for chemical or biological reaction, a detection area for determining the concentration of solution using optical source and detector from the outside, as well as outlet ports for waste-out as shown in Figure 1. All components in a MEMS fluidic device are connected with microfluidic channels. In case there is a synthesizing solution with absorption at a specific wavelength in a MEMS fluidic device, it is possible to determine the concentration by using an absorption method at specific absorption wavelength based on the Beer-Lambert law [1]<sup>1</sup>. MEMS fluidic devices are more cost-effective than conventional analysis tools and methods since expensive reagents and human power are used less and in-situ monitoring is enabled.



### Key

- 1 inlet ports
- 2 microfluidic mixer
- 3 microfluidic channel
- 4 micro-reactor
- 5 detection area
- 6 outlet port

**Figure 1 – Schematic drawing of micro components in a MEMS fluidic device (top view)**

<sup>1</sup> Numbers in square brackets refer to the Bibliography.

**SEMICONDUCTOR DEVICES –  
MICRO-ELECTROMECHANICAL DEVICES –**

**Part 48: Test method for determining solution concentration  
by optical absorption using MEMS fluidic device**

**1 Scope**

This part of IEC 62047 specifies the requirements and testing method to determine the solution concentration by optical absorption using MEMS fluidic device.

**2 Normative references**

There are no normative references in this document.