

## IEC TR 60216-7-2

Edition 2.0 2024-09

# TECHNICAL REPORT



Electrical insulating materials – Thermal endurance properties – Part 7-2: Accelerated determination of relative thermal endurance using analytical test methods (RTEA) – Results of the round robin tests to validate procedures of IEC TS 60216-7-1 by non-isothermal kinetic analysis of thermogravimetric data

INTERNATIONAL ELECTROTECHNICAL COMMISSION

ICS 19.020; 29.020; 29.035.01

ISBN 978-2-8322-9692-9

Warning! Make sure that you obtained this publication from an authorized distributor.

### **CONTENTS**

		RD	
IN		ICTION	
1	Scop	e	7
2	Norm	native references	7
3	Term	s and definitions	7
4	Test	specimens	8
5	Test	apparatus	9
	5.1	Thermogravimetric analyser (TGA)	9
	5.2	Purge gas supplied into the TGA furnace	9
6	Test	procedures	10
	6.1	General	
	6.2	Preconditioning of test samples	
	6.3	TGA tests with multiple heating rates	
	6.4	Calculation of the activation energy ( $E_a$ )	
	6.5	Determination of thermal endurance using TGA	
	6.5.1		11
	6.5.2	A 70 0	4.4
	6.5.3	material (Method A)	
_		N. P. C.	
7		d robin test results	
	7.1	TGA test results	12
	7.2	Degree of conversion correlated to the activation energy from conventional heat ageing data	12
	7.3	HICA determined by Method A and Method B	
	7.4	RTE <sub>A</sub> determined by Method A and TI <sub>A</sub> by Method B	
	7.5	Difference between RTE <sub>A</sub> and TI determined by the conventional heat	
		ageing tests	16
8	Obse	ervations from the round robin test results	
	8.1	General	18
	8.2	Sample weight variation	18
	8.3	Humidity and hydrolysis of the sample	20
	8.4	Considerations on repeatability of TGA curves	
	8.5	Baseline drift and responsiveness to heating rates of TGA	
9	Cond	lusion	25
Ar	inex A (	informative) Additional round robin studies with polybuthylene terephthalate	26
	A.1	Objectives	
	A.2	Test specimens	
	A.3	Test apparatus	
	A.4	Test procedures	
	A.5	Test results	
Ril	A.6 bliograp	Observationsbly	
ווט	onograp	····y ································	55
E!	nura 1	- Fitting curve of plots between degree of conversion and activation energy	
		ed by ISO 11358-2 [3] (example)	11

Figure 2 – Correlation between the initial sample mass of sample A and the difference of RTE <sub>A</sub> (TI <sub>A</sub> ) from TI	19
Figure 3 – Correlation between the initial sample mass of sample B and the difference of RTE <sub>A</sub> (TI <sub>A</sub> ) from TI	19
Figure 4 – Overlay charts of TGA curves in multiple heating rates in multiple laboratories (enlarged)	22
Figure 5 – Logarithm plots for activation energy calculation	23
Figure 6 – Fitting curves of degree of conversion versus activation energy by TGA	24
Figure A.1 – Effect of sample amount on E <sub>a</sub> (data provided by laboratory E)	33
Figure A.2 – Summary of factors affecting the TGA kinetic study for determination of RTE <sub>A</sub> and TI <sub>A</sub>	34
Table 1 – Heat ageing properties of the test specimens by the conventional procedure described in IEC 60216-5 [4]	9
Table 2 – Degree of conversion identical to the activation energy of the conventional heat ageing	13
Table 3 – HIC <sub>A</sub> determined by Method A and Method B for dielectric strength	13
Table 4 – HIC <sub>A</sub> determined by Method A and Method B for tensile strength	14
Table 5 – HIC <sub>A</sub> determined by Method A and Method B for impact strength	14
Table 6 – RTE $_{A}$ determined by Method A and TI $_{A}$ by Method B for dielectric strength	15
Table 7 – RTE <sub>A</sub> determined by Method A and TI <sub>A</sub> by Method B for tensile strength	15
Table 8 – RTE <sub>A</sub> determined by Method A and TI <sub>A</sub> by Method B for impact strength	16
Table 9 – Difference between RTE <sub>A</sub> or TI <sub>A</sub> , and TI for dielectric strength	16
Table 10 – Difference between RTE <sub>A</sub> or TI <sub>A</sub> , and TI for tensile strength	17
Table 11 – Difference between RTE <sub>A</sub> or TI <sub>A</sub> , and TI for impact strength	17
Table 12 – Comparison of degree of conversion with original or rerun data at 8 K/min	
Table A.1 – Heat ageing properties of the PBT test specimens by the conventional procedure in accordance with IEC 60216-5 [4]	26
Table A.2 – Degrees of conversion at the activation energy identical to that from conventional heat ageing	
Table A.3 – HIC <sub>A</sub> determined by Method A and Method B for dielectric strength	28
Table A.4 – HIC <sub>A</sub> determined by Method A and Method B for tensile strength	28
Table A.5 – HIC <sub>A</sub> determined by Method A and Method B for impact strength	29
Table A.6 – RTE $_{\hbox{A}}$ determined by Method A and TI $_{\hbox{A}}$ by Method B for dielectric strength	29
Table A.7 – RTE $_{ extsf{A}}$ determined by Method A and TI $_{ extsf{A}}$ by Method B for tensile strength	30
Table A.8 – RTE $_{ extsf{A}}$ determined by Method A and TI $_{ extsf{A}}$ by Method B for impact strength	30
Table A.9 – Difference between RTE <sub>A</sub> or TI <sub>A</sub> , and TI for dielectric strength	31
Table A.10 – Difference between RTE <sub>A</sub> or TI <sub>A</sub> , and TI for tensile strength	31
Table A.11 – Difference between RTE <sub>A</sub> or TI <sub>A</sub> , and TI for impact strength	32

#### INTERNATIONAL ELECTROTECHNICAL COMMISSION

## ELECTRICAL INSULATING MATERIALS – THERMAL ENDURANCE PROPERTIES –

Part 7-2: Accelerated determination of relative thermal endurance using analytical test methods (RTEA) – Results of the round robin tests to validate procedures of IEC TS 60216-7-1 by non-isothermal kinetic analysis of thermogravimetric data

#### **FOREWORD**

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at https://patents.iec.ch. IEC shall not be held responsible for identifying any or all such patent rights.

IEC TR 60216-7-2 has been prepared by IEC technical committee 112: Evaluation and qualification of electrical insulating materials and systems. It is a Technical Report.

This second edition cancels and replaces the first edition published in 2016. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) Annex A (informative) has been added to provide a round robin test with a different polymer type polybuthylene terephthalate (PBY) as an additional use case of the method in accordance with IEC TS 60216-7-1;
- b) Tables 3 to 11 have been corrected by adding units, and texts have been refined for more technical clarifications of the procedures and observations.

The text of this Technical Report is based on the following documents:

Draft	Report on voting
112/651/DTR	112/658/RVDTR

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this Technical Report is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at <a href="https://www.iec.ch/members\_experts/refdocs">www.iec.ch/members\_experts/refdocs</a>. The main document types developed by IEC are described in greater detail at <a href="https://www.iec.ch/publications">www.iec.ch/publications</a>.

A list of all parts in the IEC 60216 series, published under the general title *Electrical insulating materials – Thermal endurance properties*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- · withdrawn, or
- revised.

IMPORTANT – The "colour inside" logo on the cover page of this document indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

#### INTRODUCTION

IEC technical committee 112, (IEC TC 112) has been working on the development of IEC TS 60216-7-1 [1]<sup>1</sup> that considers the use of activation energy determined through thermal analytical tools plus abbreviated conventional heat ageing to determine a thermal index on a polymeric compound. At the same time, the Underwriters Laboratories Long-Term Thermal Aging Forum (UL LTTA Forum) has been discussing alternative methods that can speed up the determination of a thermal index. Members of the IEC TC 112 and of the UL LTTA Forum have made joint efforts to determine whether the Technical Specification developed by IEC TC 112 can be used to offer an alternative method of evaluating polymeric compounds for a thermal index.

Members of IEC TC 112 and the UL LTTA Forum decided to conduct a round robin test (RRT) using thermogravimetric analysis (TGA) according to ISO 11358-2 [3] on a known compound, with a known activation energy determined through conventional ageing with a view to validate the acceptability of IEC TS 60216-7-1, and to determine whether a similar thermal index can be calculated. The round robin testing was conducted with conventional TGA by multiple heating rates. However, running isothermal tests can be a follow-up of this RRT.

Numbers in square brackets refer to the Bibliography.

#### **ELECTRICAL INSULATING MATERIALS –** THERMAL ENDURANCE PROPERTIES -

Part 7-2: Accelerated determination of relative thermal endurance using analytical test methods (RTEA) - Results of the round robin tests to validate procedures of IEC TS 60216-7-1 by non-isothermal kinetic analysis of thermogravimetric data

#### 1 Scope

This part of IEC 60216 is intended to validate the procedures of IEC TS 60216-7-1 in providing a similar temperature index to conventional methods used in other parts of the IEC 60216 series.

The round robin test results do not provide statistical analysis for precision. The round robin test focuses on preliminary studies to understand the evaluation and calculation procedures, influence on apparatus, and data variance among laboratories before determination of precision.

#### 2 **Normative references**

There are no normative references in this document.