



# TECHNICAL SPECIFICATION



---

**Medical electrical equipment –  
Part 4-6: Guidance and interpretation – Voluntary guidance to help achieve  
basic safety and essential performance with regard to the possible effects of  
electromagnetic disturbances**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

---

ICS 11.040.01; 33.100.20

ISBN 978-2-8322-8809-2

**Warning! Make sure that you obtained this publication from an authorized distributor.**

# CONTENTS

FOREWORD..... 5

INTRODUCTION..... 7

1 Scope..... 8

2 Normative references ..... 8

3 Terms, definitions and abbreviated terms ..... 8

    3.1 Terms and definitions..... 8

    3.2 Abbreviated terms..... 9

4 How to use this document..... 9

5 General ..... 10

    5.1 Mitigation of effects caused by EM DISTURBANCES ..... 10

    5.2 Implementing well-proven techniques and measures for mitigating the effects that can be caused by EM DISTURBANCES ..... 10

        5.2.1 General principles ..... 10

        5.2.2 Choosing design techniques and measures from Annex A ..... 10

    5.3 Documenting the well-proven techniques and measures used for mitigating the effects that can be caused by EM DISTURBANCES ..... 13

        5.3.1 ME EQUIPMENT or ME SYSTEM..... 13

        5.3.2 Non-ME EQUIPMENT..... 14

Annex A (informative) Detailed guidance on the techniques and measures for mitigating the effects that can be caused by EM DISTURBANCES..... 15

    A.1 Techniques and measures that might be helpful in project management, planning and specification..... 15

        A.1.1 Techniques and measures for project management and planning ..... 15

        A.1.2 Techniques and measures for use when creating a design specification ..... 15

        A.1.3 Specifying EMC test standards to help ensure availability..... 16

        A.1.4 Protecting against high impact, unusual and malicious EM DISTURBANCES..... 19

    A.2 Techniques and measures that might be helpful in system design..... 20

        A.2.1 General ..... 20

        A.2.2 Separating system parts necessary for achieving BASIC SAFETY or ESSENTIAL PERFORMANCE from system parts that are not important for BASIC SAFETY or ESSENTIAL PERFORMANCE ..... 20

        A.2.3 Recording how the design specifications are achieved through design choices ..... 21

        A.2.4 Co-design electromagnetically diverse hardware/software in redundant channels..... 21

        A.2.5 System integration, installation, and commissioning ..... 24

        A.2.6 Fault detection and event data recording for later diagnosis ..... 25

        A.2.7 Improving mitigation of the effects that can be caused by EM DISTURBANCES in communication links ..... 26

        A.2.8 Synchronization and resynchronization techniques ..... 28

        A.2.9 Protection from persistent interference by monitoring retry counts ..... 28

        A.2.10 Independent detection of EM DISTURBANCES and/or EMI..... 29

        A.2.11 Protection of systems from tampering via communication links to external systems ..... 30

        A.2.12 Robust, high-specification electromagnetic mitigation ..... 31

        A.2.13 Techniques and measures to prevent RISKS being increased by virtualization of memory and PROCESS resources (e.g. a "digital twin") ..... 31

        A.2.14 Usability Engineering (Human Factors)..... 32

    A.3 Techniques and measures that might be helpful in operational design ..... 32

A.3.1	General .....	32
A.3.2	Developing appropriate operation and maintenance instructions .....	33
A.3.3	Designing appropriate maintenance techniques .....	33
A.3.4	Limiting the possibilities for operation and hence mis-operation .....	34
A.3.5	Protecting against operation errors .....	34
A.3.6	Protecting against hardware or software modifications or manipulations .....	35
A.3.7	Defensive programming techniques .....	35
A.3.8	Limited use of interrupts .....	38
A.3.9	Limited use of memory address pointer variables .....	39
A.3.10	Avoiding recursion .....	39
A.3.11	Error detection and correction for invariable memory .....	40
A.3.12	Error detection and correction techniques in redundant designs .....	41
A.3.13	Time-based error detection/correction in buses and interfaces .....	42
A.3.14	Error detection and correction for variable memory .....	43
A.3.15	Error detecting/correcting coding for ROM, RAM, buses, and interfaces .....	45
A.3.16	Error detection and correction for logic and data processing .....	45
A.3.17	Error detection and correction for electrical and electromechanical components .....	47
A.3.18	Caution when using hardware or software libraries .....	48
A.3.19	Error detection and correction for electronic components .....	48
A.3.20	Error detection/correction by monitoring program sequence (i.e. watchdogs) .....	51
A.3.21	Error detection and correction using multi-channel input/output interfaces .....	53
A.3.22	Using test patterns: static and dynamic .....	53
A.3.23	Using fiber-optic cables for signals and data communications .....	54
A.3.24	Techniques and measures for AC and DC power supplies/power converters .....	54
A.3.25	Monitoring of ventilation, cooling, and heating .....	56
A.3.26	Careful use of wireless (radio) data communications .....	57
A.3.27	Good electromagnetic engineering at every level of design .....	58
A.3.28	Design to comply with EMC test specifications as set out in A.1.3 and A.1.4 .....	58
A.3.29	De-rating of hardware components, where appropriate .....	59
A.3.30	Improve robustness of interrupts .....	59
A.4	Techniques and measures that might be helpful in implementation, integration, installation, and commissioning .....	60
A.4.1	Providing information on constraints and additional measures .....	60
A.4.2	Procuring materials, components, and products .....	61
A.4.3	Assemble/integrate according to the design for adequate mitigation of the effects that can be caused by EM DISTURBANCES .....	62
A.4.4	Install/commission according to the design for achieving adequate mitigation of the effects that can be caused by EM DISTURBANCES .....	62
A.5	Techniques and measures that might be helpful in VERIFICATION and validation (including testing) .....	62
A.5.1	Applying VERIFICATION and/or validation techniques and measures .....	62
A.5.2	VERIFICATION testing to the EMC test plan from A.1.3 and A.1.4 .....	64
A.5.3	Using non-standardized ad hoc checks or tests .....	65
A.5.4	Verifying correct installation and commissioning .....	66
A.5.5	EMC tests before and after accelerated life tests .....	67

A.6	Techniques and measures that might be helpful in operation, maintenance, repair, overhaul, refurbishment, and upgrade .....	67
A.6.1	Assessment of changes in the ELECTROMAGNETIC ENVIRONMENT.....	67
A.6.2	Assessment of continuing correct installation.....	68
A.6.3	Maintaining adequate mitigation of the effects that can be caused by EM DISTURBANCES, despite modifications or changes .....	69
A.6.4	Batch (lot) traceability .....	70
A.6.5	Component changes, new supplier, dual / alternate source.....	70
A.7	Techniques and measures that might be helpful in decommissioning .....	71
A.8	Integrating third-party items into ME EQUIPMENT or ME SYSTEMS .....	71
A.8.1	The general iterative approach .....	71
A.8.2	Suppliers' certifications and electromagnetic assessments .....	73
A.8.3	Custom-manufactured component parts.....	73
Annex B (informative)	Checklist of techniques and measures recommended for helping to achieve adequate mitigation of the effects that can be caused by EM DISTURBANCES .....	75
Bibliography	.....	89
Index of defined terms	.....	126
Figure 1	General principles for achieving mitigation of effects that can be caused by EM DISTURBANCES .....	13
Figure A.1	Choosing standard volume-manufactured subsystems and component parts .....	73
Table B.1	Checklist of techniques and measures recommended for adequate mitigation of the effects that can be caused by EM DISTURBANCES throughout the EXPECTED SERVICE LIFE .....	76

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

**MEDICAL ELECTRICAL EQUIPMENT –****Part 4-6: Guidance and interpretation – Voluntary guidance to help achieve basic safety and essential performance with regard to the possible effects of electromagnetic disturbances**

## FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

IEC TS 60601-4-6 has been prepared by subcommittee 62A: Common aspects of medical equipment, software, and systems, of IEC technical committee 62: Medical equipment, software, and systems. It is a Technical Specification.

The text of this Technical Specification is based on the following documents:

Draft	Report on voting
62A/1538/DTS	62A/1589/RVDTS

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this Technical Specification is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at [www.iec.ch/members\\_experts/refdocs](http://www.iec.ch/members_experts/refdocs). The main document types developed by IEC are described in greater detail at [www.iec.ch/publications](http://www.iec.ch/publications).

In this document, the following print types are used:

- recommendations and definitions: roman type.
- *test instructions: italic type.*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THIS TECHNICAL SPECIFICATION OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 1 includes 1.1, 1.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 1.1, 1.2 and 1.3.1 are all subclauses of Clause 1).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this document are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under [webstore.iec.ch](http://webstore.iec.ch) in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

**IMPORTANT – The "colour inside" logo on the cover page of this document indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.**

## INTRODUCTION

In 2017 it was decided to remove Annex F of IEC 60601-1-2:2014 [1]<sup>1</sup> into a separate document and provide guidance to help achieve basic safety and essential performance with regard to the possible effects of ELECTROMAGNETIC DISTURBANCE by a technical specification under the IEC 60601 series of standards.

This IEC document provides voluntary guidance on the assessment and application of techniques and measures that can help reduce the risks associated with the interfering effects of ELECTROMAGNETIC DISTURBANCES on medical equipment and medical systems.

---

<sup>1</sup> Numbers in square brackets refer to the Bibliography.

## MEDICAL ELECTRICAL EQUIPMENT –

### **Part 4-6: Guidance and interpretation – Voluntary guidance to help achieve basic safety and essential performance with regard to the possible effects of electromagnetic disturbances**

#### **1 Scope**

This document provides practical methods to help achieve BASIC SAFETY and ESSENTIAL PERFORMANCE with regard to the possible effects of EM DISTURBANCES throughout the EXPECTED SERVICE LIFE of ME EQUIPMENT or an ME SYSTEM.

These practical methods attempt to address all of the different types of errors, malfunctions or failures that can be caused by EM DISTURBANCES in ME EQUIPMENT or ME SYSTEMS.

The purpose of this document is to provide recommendations for the techniques and measures used in the design, VERIFICATION, and validation of systems, hardware, software, and firmware used in ME EQUIPMENT or ME SYSTEMS to help achieve BASIC SAFETY and ESSENTIAL PERFORMANCE with regard to the EM DISTURBANCES that could occur throughout the EXPECTED SERVICE LIFE.

#### **2 Normative references**

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60601-1, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*