

**Expression of performance of electrochemical analyzers –  
Part 2: pH value**

**CORRIGENDUM 2**

Page 18

**Table B.2**

*For buffer solution A, replace the value of the Molality by 0,05.*

*For buffer solution A, replace the value of the Mass by 12,620.*