

**Modification N° 1**  
(Décembre 1979)

à la

**Publication 523 de la CEI**  
(Première édition – 1975)

**Potentiomètres à courant continu**

**Amendment No. 1**  
(December 1979)

to

**IEC Publication 523**  
(First edition – 1975)

**Direct-current potentiometers**

**CORRIGENDA 1**

**Page 12 de la Modification N° 1**

*Au paragraphe 8.2.1, remplacer le point e) par :*

*e) en utilisant le symbole E-7, E-8 ou E-9.*

**Page 13 of Amendment No. 1**


*In Sub-clause 8.2.1, replace Item e) with :*

*e) using symbol E-7, E-8 or E-9.*

**Page 14 de la Modification N° 1**

*Remplacer la ligne introductive au-dessous des mots «Page 24» ainsi que le tableau par ce qui suit :*


*Insérer dans le tableau IV, au-dessous de E-8, le texte suivant :*

E-9	Indice de classe se référant aux erreurs exprimées en notation scientifique (par exemple $1 \times 10^{-4}$ ) lorsque l'erreur admissible est proportionnelle en partie à la valeur conventionnelle et en partie à la valeur affichée	
-----	---	---

**Page 15 of Amendment No. 1**

*Replace the line in italics under "Page 25" as well as the table by the following :*

*In Table IV, insert the following below E-8:*

E-9	Class index with errors expressed in scientific notation (e.g. $1 \times 10^{-4}$ ) when the permissible error is proportional in part to the fiducial value and in part to the dial setting	
-----	--	---